



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUL 25 2003

Mytech Technology Co., Ltd.
c/o Ms. Shu Chen Cheng
ROC Chinese-European Industrial Research Society
CEIRS
2064 Tamarin Drive
Columbus, OH 43235

Re: K030221

Trade Name: Mytech Happy Life Blood Pressure Monitor, Model HPL-300 and HPL-300A
Regulation Number: 21 CFR 870.1130
Regulation Name: Noninvasive Blood Pressure Measurement System
Regulatory Class: Class II (two)
Product Code: DXN
Dated: April 26, 2003
Received: April 30, 2003

Dear Ms. Cheng:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 – Ms. Shu Chen Cheng

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4646. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in black ink, appearing to read "Bram D. Zuckerman".

Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

MYTECH TECHNOLOGY CO., LTD.

5F.6, Alley 2, Lane 222, Lien Cheng Road,
Chung Ho City, Taipei Hsien, 235, Taiwan, ROC
Telephone: 886-2-2247 4816 Fax: 886-2-2247 7024
Email: sale@mytech.com.tw

Applicant: Mytech Technology Co., Ltd.

510(k) Number (if known): K030221


Device Name: MYTECH / HAPPY LIFE BLOOD PRESSURE MONITOR,
HPL300, HPL300A

● *Indications for use:*

The device is a noninvasive blood pressure measurement system intended to measure the systolic and diastolic blood pressures and pulse rate of an adult individual, over age 18, at home by using a non-invasive technique in which an inflatable cuff is wrapped around the arm. The cuff circumference is limited to be 7"-15".

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON
ANOTHER PAGE)

Concurrence of CDRH, office of Device Evaluation (ODE)


(Division Sign-Off)
Division of Cardiovascular Devices

510(k) Number K030221

Prescription Use _____

OR

Over — The — Counter — Use ☒

(Per 21 CFR 801.109)

(Optional Format 1-2-96)